



America

CERTIFICATE

No. U8V 101556 0004 Rev. 00

Holder of Certificate: **Communications & Power Industries
Canada Inc.**

45 River Drive
Georgetown ON L7G 2J4
CANADA

Production Facility(ies): 101556
Certification Mark:



Product: **General Medical Devices
Mobile Nano X-Ray Generator**

Model(s): **VZW2566RTDM1N-YY**
Where: 'YY' can be any alpha numeric characters representing options not affecting safety

Parameters:
Rated Input Voltage: 62-76Vdc
Rated Input Current: 0.5A nominal, 11A momentary
Rated Input Power: 8kW momentary
Protection Class: Internally Powered or Class I in end use

Tested according to:
CAN/CSA-C22.2 No. 60601-1:2014 excluding clauses 11.7 (Biocompatibility), usability (clauses 12.2 & 15.1) and 17 (EMC)
CAN/CSA-C22.2 No. 60601-2-54:2011
ANSI.AAMI ES60601-1:2005/A1:2012 excluding clauses 11.7 (Biocompatibility), usability (clauses 12.2 & 15.1) and 17 (EMC)

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169006207-001

Date, 2019-09-12

(Vivek Chandna)