




America

# CERTIFICATE

No. U10 101556 0008 Rev. 00

**Holder of Certificate:** **Communications & Power Industries  
Canada Inc.**  
45 River Drive  
Georgetown ON L7G 2J4  
CANADA

**Certification Mark:**



**Product:** **General Medical Devices  
Medical Electrical Component,  
Indico IQ High Frequency X-ray Generator**

**Tested according to:** CSA C22.2 No. 60601-1:2014/A2:2022-03  
ANSI/AAMI ES60601-1:2005/A2:2021  
CSA C22.2 No. 60601-1-6:2011/A2:2021-08  
CSA C22.2 No. 60601-2-54:2011/A2:2020-10

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. The certificate holder shall not transfer this certificate to third parties. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". For Canadian standards TÜV SÜD America Inc. is accredited by the Standards Council of Canada to ISO/IEC 17065.

**Test report no.:** 7169004676-400

**Date,** 2024-08-26

( Peter Keith )



America

# CERTIFICATE

No. U10 101556 0008 Rev. 00

## Model(s):

VZW2558FJ8YY, VZW2558FD8YY, VZW2558FC8YY, VZW2558FB8YY

Where: 'F' represents Radiographic or Fluoroscopic X-ray modes and can be R or F

'YY' represents other hardware and software options not affecting safety and can be 00 to 99 or AA to ZZ

## Parameters:

Rated voltage: 400-480Vac

Rated frequency: 50/60Hz

Rated input current: VZW2558FJ8YY: 4A Nominal, 220-170A Momentary  
 VZW2558FD8YY: 4A Nominal, 173-130A Momentary  
 VZW2558FC8YY: 3A Nominal, 136-118A Momentary  
 VZW2558FB8YY: 3A Nominal, 107-90A Momentary

Protection class: I

Degree of protection: No applied part.